

**CERTIFICATE OF COMPLIANCE**  
**RoHS Declaration**

**MANUFACTURER:** Illumina Inc.,  
5200 Illumina Way  
San Diego, CA 92122 USA

**Reference number:** 20005715  
**Product Name:** NextSeq 550Dx Sequencing System  
**Model Number:** 20005715

**DIRECTIVE 2011/65/EU OF THE EUROPEAN PARLIAMENT AND OF THE COUNCIL on the restriction of the use of certain hazardous substances in electrical and electronic equipment.**

The RoHS Directive 2011/65/EU restricts the use of certain substances including lead, mercury, cadmium, hexavalent chromium and halogenated flame retardants PBB and PBDE in electrical and electronic equipment.

This document declares that Illumina Inc. In Vitro Diagnostic (IVDs) / medical devices are in conformity with Directive 2011/65/EU of the European Parliament and of the Council of 8 June 2011 on the restriction of the use of certain hazardous substances in electrical and electronic equipment with exemptions

This declaration is based on analysis of raw materials used in the manufacturing process and suppliers' declarations.

This declaration of conformity is issued under the sole responsibility of the manufacturer.

Restricted substances referred to in Annex VI in the RoHS Directive 2011/65/EU and maximum concentration values tolerated by weight in homogeneous materials are:

Restricted hazardous substances	Abbreviation	Restricted limit (%)	Restricted limit (ppm)
Lead	Pb	0.1	1,000
Mercury	Hg	0.1	1,000
Cadmium	Cd	0.01	100
Hexavalent chromium	Cr(VI)	0.1	1,000
Polybrominated biphenyls	PBB	0.1	1,000
Polybrominated diphenylethers	PBDE	0.1	1,000

**Signature:**



**Name:** E. Joseph McMullen

**Title:** Sr. Director, Regulatory Affairs, Regulatory Affairs Department

**Date:**

14 Sept 2016