



Ref. Certif. No.

US-TUVR-013187-M1

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

CB TEST CERTIFICATE

Product

DNA Analyzer

Name and address of the applicant

Illumina, Inc.
5200 Illumina Way San Diego CA 92122
USA

Name and address of the manufacturer

Illumina, Inc.
5200 Illumina Way San Diego CA 92122
USA

Name and address of the factory

Note: When more than one factory, please report on page 2

Illumina Singapore Pte. Ltd.
#02-13/18 29 Woodlands Industrial Park E1. North Tech Lobby
3 Singapore 757716
Singapore

Ratings and principal characteristics

AC 100-240V, 50/60Hz; 300W
Class I; Laser Class 1

Trademark / Brand (if any)

illumina

Customer's Testing Facility (CTF) Stage used

N/A

Model / Type Ref.

iScan

Additional information (if necessary may also be reported on page 2)

Reissue of CB Certificate US-TUVR-13187 dated 31.10.2025 due to modification (alternate construction). See associated test report for details.

A sample of the product was tested and found to be in conformity with

IEC 61010-1:2010+A1
IEC 61010-2-081:2019
IEC 60825-1:2014
for national differences see test report

As shown in the Test Report Ref. No. which forms part of this Certificate

31470626 012

This CB Test Certificate is issued by the National Certification Body



TUV Rheinland of North America, Inc.
400 Beaver Brook Road; Boxborough, MA 01719 US
Tel: 978-266-9500
E-mail: certification@us.tuv.com
Website: www.tuv.com

Date: 2026-02-10

Signature:

James Howell
James Howell

10/061SMD 2024-12 rke-simplified