

Business Stream Products
Certification Services

Date : 03/08/2016

Illumina, Inc.
5200 Illumina Way
San Diego CA 92122
USA

Attn: Jose Briseno

Re. : CU US + Canada Certificate

Type of Equipment : Cluster Station
Model Designation : See Certificate
Certificate No. : CU 72160528 0001
File No. : 31690679 002
Engineer/Contact : Jorge Rubio
Standard(s) : UL 61010-1:2012
UL 61010-2-010:2015
CAN/CSA-C22.2 NO. 61010-1-12
CAN/CSA-C22.2 NO. 61010-2-010:15

Dear Mr. Briseno,


The above referenced technical equipment has been tested and was found to be in compliance with the listed test requirement(s). Enclosed, please find the TUV Rheinland approval document No. CU 72160528 0001. It authorizes you to label the listed product(s) with the TUV Rheinland Mark identified in the approval document. For compliance, the Test Mark must be on the approved unit.

Your product is subject to regular factory follow-up inspections as well as annual certificate and factory registration fees.

In using the TUV Rheinland Mark you are obligated to comply with the TUV Rheinland of North America Service Agreement.

If we can be of any further assistance to you, please do not hesitate to contact us.

Sincerely yours,
Certification Body


Dipl.-Ing. M. Glagla
QA Certification Officer

TUV Rheinland
of North America, Inc.
North American Headquarters

12 Commerce Road
Newtown, CT 06470

Tel 203-426-0888
Fax 203-426-4009
Toll Free TUV-RHEINLAND
Mail info@tuv.com
Web www.tuv.com

Member of
TUV Rheinland Group

Enclosure

Certificate



Certificate no.

CU 72160528 01

License Holder:

Illumina, Inc.
5200 Illumina Way
San Diego CA 92122
USA

Manufacturing Plant:

VM Services Inc.
6723 Mowry Ave
Newark NJ 94560
USA

Test report no.: USA-JR 31690679 002

Client Reference: Jose Briseno

Tested to: UL 61010-1:2012
UL 61010-2-010:2015
CAN/CSA-C22.2 NO. 61010-1-12
CAN/CSA-C22.2 NO. 61010-2-010:15

Certified Product: Cluster Station

License Fee - Units

Model Designation: cBot, cBot 2

7

Rated Voltage: AC 100-240V, 50/60Hz
Rated Power: 500W
Protection Class: I

Appendix: 1, 1-6

7

Licensed Test mark:

Date of Issue

(day/mo/yr)

08/03/2016

