



Declaration of Conformity

Illumina, Inc. hereby declares under its sole responsibility that the product(s) listed are in conformity to the EMC Directive [2014/30/EU], Low Voltage Directive [2014/35/EU], RED Directive [2014/53/EU] and RoHS Directive [2011/65/EU] as amended by Commission Delegated Directive (EU) 2015/863.

MANUFACTURER:
ADDRESS:

Illumina, Inc
5200 Illumina Way
San Diego, CA 92122, USA

FACTORY LOCATION:
25841 Industrial Blvd.
Hayward, CA 94545, USA

PRODUCT TYPE:
MODEL:
CE MARK AFFIXED:

Next Generation Sequencer
MiSeq, MiSeq FGx
2011

AUTHORIZED EU REPRESENTATIVE:
Illumina Netherlands B. V.
Steenoven 19
5626 DK Eindhoven
The Netherlands

The construction of the product is in compliance with the following harmonized and/or consensus standards.

| | |
|--|--|
| EN 61010-1:2010+A1:2019 | <i>Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 1: General Requirements</i> |
| EN 61010-2-081:2019 | <i>Particular requirements for automatic and semi-automatic laboratory equipment for analysis and other purposes</i> |
| EN IEC 61326-1:2021 (Class A) | <i>Electrical equipment for the measurement, control and Laboratory use – EMC Requirements Part1, Class A</i> |
| EN 55011:2011: 2016+A1:2017+A11:2020+A2:2021 | <i>Industrial, scientific and medical equipment - Radio-frequency disturbance characteristics - Limits and methods of measurement</i> |
| EN 55032:2012/A:2013 | <i>Information technology equipment - Radio disturbance characteristics - Limits and methods of measurement</i> |
| EN 301 489-1 V2.2.3 | <i>EMC Standard for radio equipment and services; Part 1: Common technical requirements; Harmonized Standard covering the essential requirements of article 6 of Directive 2014/30/EU</i> |
| EN 301 489-3 V2.1.1 | <i>EMC standard for radio equipment and services; Part 3: Specific conditions for Short-Range Devices (SRD) operating on frequencies between 9 kHz and 246 GHz</i> |
| EN 62311:2008 | <i>Assessment of electronic and electrical equipment related to human exposure restrictions for electromagnetic fields (0 Hz to 300 GHz)</i> |
| EN 300 330 V2.1.1 | <i>"Short Range Devices (SRD); Radio equipment in the frequency range 9 kHz to 25 MHz and inductive loop systems in the frequency range 9 kHz to 30 MHz; Harmonized Standard covering the essential requirements of article 3.2 of Directive 2014/53/EU "</i> |
| EN/IEC 61000-3-2:2019+A1:2020 | <i>Electromagnetic compatibility (EMC) - Part 3-2: Limits - Limits for harmonic current emissions (equipment input current ≤ 16 A per phase)</i> |
| EN/IEC 61000-3-3:2013+A2:2021 | <i>Electromagnetic compatibility (EMC) - Part 3-3: Limits - Limitation of voltage changes, voltage fluctuations and flicker in public low-voltage supply systems, for equipment with rated current ≤16 A per phase and not subject to conditional connection</i> |
| EN 63000:2018 | <i>Technical documentation for the assessment of electrical and electronic products with respect to the restriction of hazardous substances</i> |

Illumina declares the product listed above is in compliance with RoHS Directive 2011/65/EU, as amended by (EU) 2015/863 on the restriction of the use of certain hazardous substances in electrical and electronic equipment.

This declaration is based on analysis of raw materials used in the manufacturing process and supplier's declarations.

| | |
|---------------------------------------|---|
| Lead (0,1%) | Polybrominated diphenylethers (PBDE) (0,1%) |
| Mercury (0,1%) | Bis(2-Ethylhexyl) phthalate (DEHP) (0,1%) |
| Cadmium (0,01%) | Benzyl butyl phthalate (BBP) (0,1R%) |
| Hexavalent chromium (0,1%) | Dibutyl phthalate (DBP) (0,1%) |
| Polybrominated biphenyls (PBB) (0,1%) | Diisobutyl phthalate (DIBP) (0,1%) |

Annex III exemptions are applied.

Authorized by:


Ralph Jones
VP, Instrument Manufacturing

Electronically signed by: Ralph Jones
Reason: Approver
Date: Dec 9, 2025 14:21:59 PST

Date