

Date: 2025-12-11

Illumina, Inc.
5200 Illumina Way
San Diego CA 92122
USA

Ref: CU US + Canada Certificate

Type of Equipment : Genetic Sequencer
Certificate No. : CU 72504916 0001
Report No. : 31181842 057
Engineer/Contact : Edalyn Canete
Standards : UL 61010-1:2012 R7.19
UL 61010-2-081:2019
CAN/CSA-C22.2 NO. 61010-1-12 + GI1 + GI2 (R2017) + A1
CSA C22.2 No. 61010-2-081:19

Dear Madame or Sir,

The above referenced technical equipment has been tested and was found to be in compliance with the listed test requirement(s). Enclosed, please find the TUV Rheinland approval document No. CU 72504916 0001.

It authorizes you to label the listed product(s) with the TUV Rheinland Mark identified in the approval document. For compliance, the Test Mark must be on the approved unit.

Your product is subject to regular factory follow-up inspections as well as annual certificate and factory registration fees.

In using the TUV Rheinland Mark you are obligated to comply with the TUV Rheinland of North America Service Agreement.

If we can be of any further assistance to you, please do not hesitate to contact us.

With kind regards,

Certification Body

Dipl.-Ing. M. Glagla

Enclosure

Certificate

Certificate no.

CU 72504916 0001

License Holder:

Illumina, Inc.
5200 Illumina Way
San Diego CA 92122
USA

Manufacturing Plant:

Illumina Inc.
25841 Industrial Blvd
Hayward CA 94545
USA

Report Number: 31181842 057

Client Reference: Carol Escano

Certification acc. to: UL 61010-1:2012 R7.19

UL 61010-2-081:2019

CAN/CSA-C22.2 NO. 61010-1-12 + GI1 + GI2 (R2017) + A1

CSA C22.2 No. 61010-2-081:19

Product Information

Certified Product: Genetic Sequencer

Model Designation: MiSeq, MiSeqFGx

Technical Data: Rated Current: AC 100-240V, 50/60Hz

Rated Power: 400W

Protection Class: I

Remarks: Replaces Certificate CU72300009.

Appendix: 1, 1-7



Date of issue: 2025-12-11
(yr/mo/day)

TUV Rheinland of North America, Inc.
400 Beaver Brook Rd, Boxborough, MA 01719
Tel +1 (978) 266 9500, Fax +1 (978) 266-9992

www.tuv.com

TÜV Rheinland®