



Ref. Certif. No.

US-TUVR-013287

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT
(IECEE) CB SCHEME

CB TEST CERTIFICATE

Product

Genetic Sequencer

Name and address of the applicant

Illumina, Inc.
5200 Illumina Way San Diego CA 92122
USA

Name and address of the manufacturer

Illumina, Inc.
5200 Illumina Way San Diego CA 92122
USA

Name and address of the factory

Note: When more than one factory, please report on page 2

Illumina Inc.
25841 Industrial Blvd Hayward CA 94545
USA

Ratings and principal characteristics

AC 100-240V, 50/60Hz; 400W
Class I

Trademark / Brand (if any)

illumina

Customer's Testing Facility (CTF) Stage used

N/A

Model / Type Ref.

MiSeqDx

Additional information (if necessary may also be reported on page 2)

Replaces CB Certificate DE 2038793 due to change of NCB.

A sample of the product was tested and found to be in conformity with

IEC 61010-1:2010+A1
IEC 61010-2-081:2019
IEC 61010-2-101:2018
for national differences see test report

As shown in the Test Report Ref. No. which forms part of this Certificate

31181842 056

This CB Test Certificate is issued by the National Certification Body



TUV Rheinland of North America, Inc.
400 Beaver Brook Road; Boxborough, MA 01719 US
Tel: 978-266-9500
E-mail: certification@us.tuv.com
Website: www.tuv.com

Date: 2025-12-02

Signature: Dipl.-Ing. M. Glagla