



Ref. Certif. No.

US-TUVR-010266-A3

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

CB TEST CERTIFICATE

Product

Next Generation Sequencer

Name and address of the applicant

Illumina, Inc.
5200 Illumina Way
San Diego CA 92122, USA

Name and address of the manufacturer

Illumina, Inc.
5200 Illumina Way
San Diego CA 92122, USA

Name and address of the factory

(See appendix for factories information)

Ratings and principal characteristics

AC 200-240V, 50/60Hz; 15A (2500W); Class I; Laser Class 1

Trademark (if any)

Illumina

Customer's Testing Facility (CTF) Stage used

Model / Type Ref.

NovaSeq 6000

Additional information (if necessary may also be reported on page 2)

Re-issuance of CB Certificate US-TUVR-10266 due to 3rd modification (alternate construction and components). See associated test report for details.

A sample of the product was tested and found to be in conformity with

IEC 61010-1:2010
IEC 61010-2-010:2014
IEC 61010-2-081:2015
IEC 60825-1:2014
for national deviations see test report

As shown in the Test Report Ref. No. which forms part of this Certificate


31790423 007

This CB Test Certificate is issued by the National Certification Body



TUV Rheinland of North America, Inc.
12 Commerce Road
Newtown, CT 06470 - USA

Date: 24.08.2018

Signature: 
Marten Glagla

1. Illumina, Inc.
25861 Industrial Blvd.
Hayward CA 94545
USA

2. VM Services Inc.
6723 Mowry Ave
Newark CA 94560
USA

Additional information (if necessary)
Information complémentaire (si nécessaire)

Date:

24.08.2018

Signature:



Martin Giagla